



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/970,390	10/02/2001	Daniel R. Kürz	MICRU : 58614	1995
24201	7590	06/14/2005	EXAMINER	
FULWIDER PATTON LEE & UTECHT, LLP HOWARD HUGHES CENTER 6060 CENTER DRIVE TENTH FLOOR LOS ANGELES, CA 90045			BAXTER, JESSICA R	
			ART UNIT	PAPER NUMBER
			3731	

DATE MAILED: 06/14/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

SP

Office Action Summary	Application No. 09/970,390	Applicant(s) KURZ ET AL.	
	Examiner Jessica R. Baxter	Art Unit 3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 11 April 2005.
 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 54, 57 and 59-65 is/are pending in the application.
 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
 5) ☐ Claim(s) _____ is/are allowed.
 6) ☒ Claim(s) 54, 57 and 59-65 is/are rejected.
 7) ☐ Claim(s) _____ is/are objected to.
 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 a) ☐ All b) ☐ Some * c) ☐ None of:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 54, 57, 61, 62, 63, 64 and 65 are rejected under 35 U.S.C. 102(b) as being anticipated by U.S. Patent No. 5,639,277 to Mariant et al.

Mariant discloses an occlusive device comprising a vasoocclusive coil having a primary coil configuration with two-dimensional J-shaped loops (FIG. 1, FIG. 3, FIG. 5) at the proximal and distal ends of the coil, said vasoocclusive coil including a radiopaque strand (Column 4 lines 3-6).

Claim Rejections - 35 USC § 103

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. Claims 59 and 60 are rejected under 35 U.S.C. 103(a) as being unpatentable over Mariant et al. '277 in view of U.S. Patent No. 5,733,329 to Wallace et al.

Mariant discloses the claimed invention except for the diameter of the helical loops. Wallace teaches that a 2 mm device may be used in vasoocclusive applications depending on the particular site within the human body (Column 5 lines 5-15). It would have been obvious to

one having ordinary skill in the art at the time the invention was made to provide the device of Mariant with the diameter of Wallace in order to place the device of Mariant in an appropriately sized vessel.

Response to Arguments

5. Applicant's arguments filed April 11, 2005 have been fully considered but they are not persuasive.

6. Applicant argues that Mariant et al. ' does not disclose a primary coil configuration formed about a substantially single axis. However, the primary coil configuration is formed about a substantially single axis. The secondary configuration of the coils is formed about multiple axes. The two dimensional J-loop will be seen at the end of any of the primary coil configurations since the end of a coil ends in a J shape as a portion of the looping of the coil. Applicant further argues that the J-shaped loop is not part of the secondary configuration, but rather is claimed as part of the primary coil configuration. The claim is interpreted as the vasoocclusive coil having "a two-dimensional J-shaped loop at at least one end." This J-shape is clearly seen in the Figures (see FIG. 1) of Mariant et al. '277. Therefore, the rejection over Mariant et al. '277 is proper.

Conclusion

7. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be

Art Unit: 3731

calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jessica R. Baxter whose telephone number is 571-272-4691. The examiner can normally be reached on M-F 8:30AM - 5:00PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anhtuan T. Nguyen can be reached on 571-272-4963. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Jessica R Baxter
Examiner
Art Unit 3731

JB
rb

ANHTUAN T. NGUYEN
ANHTUAN T. NGUYEN
SUPERVISORY PATENT EXAMINER

6/9/05